



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/662,877	09/16/2003	Richard J. Whitbourne	32286-191984	1150
26694	7590	09/18/2006	EXAMINER	
VENABLE LLP P.O. BOX 34385 WASHINGTON, DC 20043-9998			HO, UYEN T	
			ART UNIT	PAPER NUMBER
			3731	

DATE MAILED: 09/18/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary	Application No. 10/662,877	Applicant(s) WHITBOURNE ET AL.	
	Examiner (Jackie) Tan-Uyen T. Ho	Art Unit 3731	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 16 September 2003.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-52 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-52 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Response to Arguments

1. Applicant's arguments have been considered but are moot in view of the new ground(s) of rejection.

Claim Rejections - 35 USC § 102

2. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

3. Claims 1, 3-13, 15-29, 31-39, 42, 46, 47-52, rejected under 35 U.S.C. 102(e) as being anticipated by Pacetti et al. (6,663,662). Pacetti et al. disclose a stent having a coating comprising a primer layer as claimed (col. 4, line 39 to col. 8, line 44), a biologically active agent as claimed (col. 11, lines 1-65), an outer most drug reservoir comprising layers 34 and 28 (col. 8, line 45 to col. 15, line 46), image enhancing material (col. 14, lines 15-36), the thickness as claimed (col. 18).

The primer layer comprising two or more polymers (col. 16, line 46-col. 17, line 62).

The primer layer is a single layer (32)

The primer layer polymers being distinct from the drug reservoir layer polymers. First, the polymers of drug reservoir layer contain active ingredient and the primer layer does not. Second the polymers of drug reservoir layer comprising particles and the primer layer does not. Third, Pacetti et al. disclose the composition of an primer layer in col. 4, line 39 to col. 8, line 44 some are different from the composition of forming the active ingredient coating layer disclosed in col. 8, line 45 to col. 15, line 46).

The claim limitations does not exclude the drug reservoir from having a diffusion layer.

Claim Rejections - 35 USC § 103

4. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

5. Claims 2, 14, 30 are rejected under 35 U.S.C. 103(a) as being unpatentable over Pacetti et al. (6,663,662). Although, Pacetti et al. do not disclose an intermediate layer between the primer layer and drug reservoir layer, it is well known in the art to have an intermediate layer as claimed in order to enhance the attachment between polymer layers. Therefore, it would have been obvious to one having ordinary skill in the art at the time the invention was made to employ an intermediate layer between the primer

layer and drug reservoir layer of Pacetti et al.'s stent in order to enhance the attachment between layers.

6. Regarding to claims 40, 41, 43-45, although Pacetti et al. do not disclose the type of the ethylene acrylic acid copolymer or the polyurethane as claimed, these types of materials as claimed are well known in the art. Therefore, it would have been obvious to one having ordinary skill in the art at the time the invention was made to the primer layer and drug reservoir layer of Pacetti et al. the specific type of ethylene acrylic acid copolymer and the polyurethane as claimed. Doing so would amount to mere substitution of one material for another within the same art that would perform equally well in the Pacetti et al. stent.

Conclusion

7. Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of

Art Unit: 3731

the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

8. Any inquiry concerning this communication or earlier communications from the examiner should be directed to (Jackie) Tan-Uyen T. Ho whose telephone number is 571-272-4696. The examiner can normally be reached on MULTIFLEX Mon. to Sat..

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, ANHTUAN NGUYEN can be reached on 571-272-4963. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



(Jackie) Tan-Uyen T. Ho
Primary Examiner
Art Unit 3731

September 13, 2006